

Laucks

Testing Laboratories, Inc.

940 South Harney St., Seattle, Washington 98108 (206) 767-5060 FAX 767-5063

Chemistry, Microbiology, and Technical Services



Alaskan Copper Works
P. O. Box 3546
Seattle, WA 98124


Date Received: 10/23/89
Date Reported: 11/01/89
Work Order: 89-10-177
Category: 1023632

Attn: Raphael Acholonu

Work ID: Metro Waste Discharge
P O # : M12179

Test	Units	891018 Metro Split Grab 10/19/89
Chromium (Method 218.1)	mg/L	0.63
Copper (Method 220.1)	mg/L	0.24
Nickel (Method 249.1)	mg/L	1.1
Zinc (Method 289.1)	mg/L	0.080

Sample was further identified as: 10-18-3200-M12179.

Certified By: 

Charter Member American Council of Independent Laboratories

AKC-0015118

Laucks

Testing Laboratories, Inc.

940 South Harney St., Seattle, Washington 98108 (206) 767-5060 FAX 767-5063

Chemistry, Microbiology, and Technical Services

REPORT ON WORK ORDER 8910177 PREPARATION BLANKS

Test : Zinc (Method 289.1)
Blank Name : B1025AA01 Preparation Date: 10/25/89
Conc Found : 0.002 U Control Limit : 0.004
Units : mg/L

This blank and comments, if any, apply to the following sample(s):
1

Test : Copper (Method 220.1)
Blank Name : B1025AA01 Preparation Date: 10/25/89
Conc Found : 0.002 U Control Limit : 0.004
Units : mg/L

This blank and comments, if any, apply to the following sample(s):
1

Test : Nickel by FAA in water
Blank Name : B1025AA01 Preparation Date: 10/25/89
Conc Found : 0.010 U Control Limit : 0.020
Units : mg/L

This blank and comments, if any, apply to the following sample(s):
1

Test : Chromium (Method 218.1)
Blank Name : B1025AA01 Preparation Date: 10/25/89
Conc Found : 0.005 U Control Limit : 0.010
Units : mg/L

This blank and comments, if any, apply to the following sample(s):
1

Charter Member American Council of Independent Laboratories

AKC-0015119

Laucks

Testing Laboratories, Inc.

940 South Harney St., Seattle, Washington 98108 (206) 767-5060 FAX 767-5063

Chemistry, Microbiology, and Technical Services

REPORT ON WORK ORDER 8910177 PREPARATION BLANKS

* = outside control limits
U = analyte not detected

Charter Member American Council of Independent Laboratories

AKC-0015120